

ATP Test Report Sum

Test Name	Device Under Test
Max Power vs Frequency	M801010016
Power Linearity vs Frequency	M801010016
Power Resolution	M801010016
Frequency Accuracy	M801010016
Harmonic Distortion	M801010016
Sub-Harmonic Distortion	M801010016
Spurious In-Band	M801010016
Amplitude Modulation Depth	M801010016
Amplitude Modulation Frequency	M801010016
Frequency Modulation Frequency	M801010016
Phase Modulation Frequency	M801010016
SSB Phase Noise	M801010016

Summary – Device M801010016, LSX 8GHz

Test Equipment	Pass	Fail
Keysight Technologies N9000B (S/N: MY57103246)	441	0
Keysight Technologies N9000B (S/N: MY57103246)	268	0
Berkeley Nucleonics Corp BNC12144L (S/N: B10224)	409	0
Berkeley Nucleonics Corp BNC12144L (S/N: B10224)	27	0
Berkeley Nucleonics Corp BNC12144L (S/N: B10224)	29	0
Keysight Technologies N9000B (S/N: MY57103246)	152	0
Keysight Technologies N9000B (S/N: MY57103246)	160	0
Keysight Technologies N9000B (S/N: MY57103246)	11	0
Keysight Technologies N9000B (S/N: MY57103246)	102	0
Keysight Technologies N9000B (S/N: MY57103246)	131	0
Keysight Technologies N9000B (S/N: MY57103246)	128	0
Rohde & Schwarz FSPN-26 (S/N: 1322.8003K25/101335)	6	0